<https://helda.helsinki.fi>

A body-sized phantom for evaluation of diffusion-weighted MRI data using conventional, readout-segmented, and zoomed echo-planar sequences

Ihalainen, Toni

2016

Ihalainen , T , Kuusela , L , Soikkeli , M , Lantto , E , Ovissi , A & Sipilä , O 2016 , ' A body-sized phantom for evaluation of diffusion-weighted MRI data using conventional, readout-segmented, and zoomed echo-planar sequences ' , Acta Radiologica , vol. 57 , no. 8 , pp. 947 - 954 . <https://doi.org/10.1177/0284185115613652>

<http://hdl.handle.net/10138/223813>

<https://doi.org/10.1177/0284185115613652>

publishedVersion

Downloaded from Helda, University of Helsinki institutional repository.

This is an electronic reprint of the original article.

This reprint may differ from the original in pagination and typographic detail.

Please cite the original version.

A body-sized phantom for evaluation of diffusion-weighted MRI data using conventional, readout-segmented, and zoomed echo-planar sequences

Toni Ihalainen^{1,2}, Linda Kuusela^{1,2}, Maiju Soikkeli³, Eila Lantto¹, Ali Ovissi¹ and Outi Sipilä⁴

Abstract

Background: Abdominal diffusion-weighted imaging (DWI) has been rapidly increasing during the last few years. For the evaluation of new DWI techniques, the development of suitable phantoms and quality assurance methods is important.

Purpose: To construct a body-diameter phantom for abdominal DWI and study the impact of different acquisition options on image quality.

Material and Methods: A phantom with a diameter of 31 cm and a volume of 26 L was constructed, containing four samples representing a clinically relevant range of apparent diffusion coefficient (ADC) values. Measurements were carried out on 1.5T and 3.0T MRI systems using conventional echo-planar imaging (EPI), readout-segmented EPI, and zoomed EPI (3.0T) sequences. The effects of parallel imaging, coil intensity normalization, and patient-specific B₁ shim (3.0T) were also examined. ADC values and signal-to-noise ratios of the samples were measured, and the level of artifacts was visually evaluated.

Results: The agreement of ADC values between different acquisition options was generally good, but higher values (by $0.07 \times 10^{-3} \text{ mm}^2/\text{s}$ on the average) with readout-segmented EPI as well as ADC variations of approximately $0.1 \times 10^{-3} \text{ mm}^2/\text{s}$ in slice direction were observed. The image artifacts were reduced by using patient-specific B₁ shim, readout-segmented EPI, or zoomed EPI.

Conclusion: The body-sized phantom demonstrated well the expected image artifacts in DWI with large field of view. The use of patient-specific B₁ shim, readout-segmented EPI, or zoomed EPI improved image quality of DWI in this study.

Keywords

Magnetic resonance (MR) diffusion/perfusion, imaging sequences, quality assurance/quality control (QA/QC)

Date received: 20 March 2015; accepted: 25 September 2015

Introduction

In magnetic resonance imaging (MRI), the importance of diffusion-weighted imaging (DWI) has been substantially growing during the last few years. Having long been an essential tool for neuroradiologists, the center of mass of clinical DWI has now moved closer to the center of mass of the human body – the abdomen. An important consensus paper by Padhani et al. (1) described the possibilities and partially unsolved problems of DWI in oncology. The standardization of imaging protocols and the comparability of apparent diffusion coefficient (ADC) values between different centers were in the focus of discussion.

One of the challenges stated by the consensus paper was the development of suitable phantoms and quality

¹HUS Medical Imaging Center, Radiology, University of Helsinki and Helsinki University Hospital, Helsinki, Finland

²Department of Physics, University of Helsinki, Helsinki, Finland

³Department of Chemistry, University of Helsinki, Helsinki, Finland

⁴HUS Medical Imaging Center, Clinical Physiology and Nuclear Medicine, University of Helsinki and Helsinki University Hospital, Helsinki, Finland

Corresponding author:

Toni Ihalainen, HUS Medical Imaging Center, PO Box 340, 00029 HUS, Finland.

Email: toni.ihalainen@hus.fi

assurance methods for DWI. Delakis et al. have developed and applied a quality assurance protocol for diffusion MRI studies (2). Chenevert et al. have published a study of an ice-water phantom (3), followed by a repeatability and reproducibility study (4). Lavdas et al. have introduced a DWI phantom using different concentrations of agarose, sucrose, and nickel, and used it for comparison of 1.5T and 3.0T MRI systems (5,6). Other studies have investigated suitable materials for DWI phantoms (7–9). Miquel et al. have studied the repeatability of abdominal DWI, with phantom and *in vivo* (10). None of the mentioned phantoms have represented the dimensions of the abdominal area of a normal-sized or large patient.

Abdominal imaging at 3.0T has been problematic because of the signal inhomogeneity due to standing waves in the body area, leading to inhomogeneous distribution of the B_1 field and hence incomplete excitation of the protons (11,12). Parallel transmission has been developed to overcome these problems (13). Parallel transmission allows also the shimming of the B_1 field and the excitation of a selective volume. This approach can be used in zoomed diffusion-weighted echo-planar imaging (EPI) that utilizes smaller number of k-space lines leading to fewer artifacts in the final image (14,15). In addition, Porter et al. (16) have published a method for readout-segmented EPI, which divides the k-space into several segments in readout direction. This method substantially reduces the artifacts caused by phase errors in conventional EPI sequences.

The objectives of this study were: (i) to construct a body-sized phantom suitable for the image quality evaluation of abdominal DWI; and (ii) to investigate the uniformity of the ADC values, signal-to-noise ratios (SNR), and the level of artifacts with different field strengths and acquisition options, including parallel imaging, coil intensity normalization function, patient-specific B_1 shim, readout-segmented EPI, and zoomed EPI.

Material and Methods

Phantom construction

Agarose, sucrose, NaCl, and NaN_3 were mixed with distilled water and the mixture was stirred at 120°C until all solid material was dissolved. The mixture was cooled to 80°C and $\text{Ni}(\text{NO}_3)_2 \cdot 6 \text{H}_2\text{O}$ was added. The mixture was then poured into a 1 L plastic sample bottle (diameter, 8 cm) and cooled to room temperature. Four samples were prepared according to the method described above. The compositions of the samples are described in Table 1. The concentrations were selected to produce a clinically relevant range of ADC

Table 1. The compositions of the samples inside the phantom.

Sample	Agarose (g/L)	Sucrose (g/L)	Nickel (g/L)	NaCl (g/L)	NaN_3 (g/L)
1	8	380	0	9	0.3
2	8	140	0.17	9	0.3
3	8	0	0.52	9	0.3
4	12	0	0.52	9	0.3

Fig. 1. The numbering of the samples 1–4 (S1–S4) and examples of the regions of interest (ROI) used in the image analysis, superimposed on an axial T1W image (2D FLASH sequence: TR, 169 ms; TE, 4.76 ms; flip angle, 70°) of the phantom acquired with a 1.5T system.

values (approximately 0.8, 1.3, 1.9, and $>2.0 \times 10^{-3} \text{ mm}^2/\text{s}$), as demonstrated by Lavdas et al. (5). The sample bottles were placed in a watertight polyethylene barrel (diameter, 31 cm; height, 43 cm; volume, 26 L). The external dimensions were selected to reflect the abdominal region of a large patient. A plastic holder was constructed to hold the bottles in a correct position inside the barrel, which was then filled with background material (sucrose 400 g/L, NaCl 9 g/L). An axial T1-weighted (T1W) image of the phantom with sample numbering is presented in Fig. 1.

Image acquisition

The images were acquired with Siemens Magnetom Aera 1.5T and Siemens Magnetom Skyra 3.0T (Siemens Healthcare, Erlangen, Germany). Both systems were completely installed in a clinical site but not yet taken into clinical use. The phantom was brought to the imaging room one day before the measurements to let its

temperature stabilize at the room temperature. Before and after the measurements, the temperature of the room was measured with a Protimeter Hygromaster (GE Measurement & Control, Billerica, MA, USA). The temperature changed from 22.1°C to 22.6°C in the 1.5T room and from 19.3°C to 19.1°C in the 3.0T room during the measurements. On both MRI systems, the phantom was positioned on a 32-channel spine coil and an 18-channel body coil was secured on top of it (Fig. 2).

The following series were acquired:

- (i) a conventional abdominal DWI EPI series;
- (ii) series no. 1 without parallel imaging, increasing the TR value to 6800 ms (1.5T) or 7000 ms (3.0T) and TE values to 111 ms;
- (iii) series no. 1 with coil intensity normalization (pre-scan normalize);

Fig. 2. Positioning of the phantom.

- (iv) readout-segmented EPI (syngo RESOLVE, Siemens Healthcare);
- (v) zoomed EPI (syngo ZOOMit, Siemens Healthcare; only 3.0T) on four separate field of view (FOV) locations, each covering one sample in the phantom; and
- (vi) series no. 1 with patient-specific B₁ shim option (only 3.0T).

The imaging parameters for both systems are presented in Table 2. To allow noise evaluation with the difference method (17), two consecutive acquisitions were carried out for each series. Trace-weighted images with different b-values and ADC maps were calculated by the MRI system.

Measurement of the ADC values

The mean ADC values and standard deviations (SD) were measured from the ADC map of the central slice of each acquired series with ImageJ version 1.44p (18). Two regions of interest (ROI) were placed on each of the four samples. An oval ROI ($1200 \pm 24 \text{ mm}^2$) was placed on an artifact-free region. A circular ROI covered 75% of the sample ($3800 \pm 80 \text{ mm}^2$). Examples of the ROIs are presented in Fig. 1. For inter-system comparisons, the ADC values were corrected for the average temperature difference of 3.2°C between the two examination rooms, using the correction factor of 2.4%/°C (19). To check the ADC uniformity in slice direction, the ADC values from all slices of conventional, readout-segmented EPI, and zoomed EPI series were measured for sample 1 with the 1200 mm² ROI.

Table 2. Key imaging parameters of the conventional, readout-segmented-EPI and zoomed EPI series.

	Conventional EPI 1.5T	Conventional EPI 3.0T	Readout-segmented EPI 1.5T	Readout-segmented EPI 3.0T	Zoomed EPI 3.0T
b-values (s/mm ²)	0 / 400 / 800	0 / 400 / 800	0 / 800	0 / 800	0 / 400 / 800
Averages for each b-value	2 / 2 / 3	2 / 3 / 4	1 / 3	1 / 3	2 / 5 / 12
TR / TE (ms)	5500 / 71	5800 / 71	7300 / 83	7500 / 79	4900 / 78
Slices (n)	25	25	25	25	20
Slice thickness / gap (mm)	6 / 1.2	6 / 1.2	6 / 1.2	6 / 1.2	6 / 1.2
FOV (mm)	350 × 350	350 × 350	350 × 350	350 × 350	150 / 93.3
Acquisition matrix	192 × 154	192 × 154	192 × 154	192 × 154	90 × 56
Partial Fourier	7 / 8	7 / 8	Off	Off	Off
Receiver bandwidth (Hz/pixel)	1736	1628	592	651	1502
Acquisition time (min:s)	1:50	2:42	8:47	9:02	4:36
Parallel imaging (factor)	GRAPPA (2)	GRAPPA (2)	Off	Off	Off
Prescan normalize	Off	Off	Off	Off	Off
Fat suppression	SPAIR	SPAIR	SPAIR	SPAIR	SPAIR

EPI, echo planar imaging; FOV, field of view; GRAPPA, generalized autocalibrating partially parallel acquisitions; SPAIR, spectral attenuated inversion recovery; TE, echo time; TR, repetition time

SNR measurements

Noise images, i.e. subtraction images from two consecutive acquisitions were calculated with ImageJ for trace-weighted images ($b = 800 \text{ s/mm}^2$). Before subtraction, an arbitrary fixed value of 500 was added to the pixel values of the images of the first acquisition to avoid problems with negative pixel values. SNR values were calculated from the central-slice image for each of the four samples by:

$$SNR_n = \frac{S_n}{N_n/\sqrt{2}}$$

where S_n is the signal of the sample n , i.e. mean value of the respective ROI in the signal image, and N_n is the noise of the sample n , i.e. SD of the respective ROI in the noise image. The same oval 1200 mm^2 ROIs as in the ADC evaluation were used.

Assessment of the artifacts

A visual assessment of artifacts was carried out in consensus by a medical physicist and an MRI scientist with 10 and 7 years of experience in MRI quality assurance, respectively. The assessment was performed for two trace-weighted central-slice images ($b = 0 \text{ s/mm}^2$, $b = 800 \text{ s/mm}^2$) and the ADC map of each acquisition option, at both field strengths when applicable. To better visualize the artifacts, edge detection images were calculated for all the assessed images with Matlab 2014a (The MathWorks, Inc., Natick, MA, USA) using the Sobel filter. They represented image intensity gradient magnitude, and they were scaled utilizing the mean intensity of the original slice.

Each image to be evaluated was opened side by side with an edge detection image calculated from it. Each sample 1–4 was given an image quality score of 0–3 (0, no artifact or obtrusive local noise; 1, small or mild single artifact or some obtrusive noise; 2, substantial artifact or noise, yet enabling a reasonable ADC measurement or reading of the image; 3, artifact or noise that notably affects the ADC measurement or reading of the image). All scores of each technique at both field strengths were summed and a relative artifact level index was calculated for each acquisition option AO by:

$$\begin{aligned} \text{Artifact level index (AO)} \\ = \frac{\text{Sum of scores (AO)}}{\text{Sum of scores (conventional EPI)}} \times 100 \end{aligned}$$

The artifact level indices for patient-specific B_1 shim and zoomed EPI options were calculated based on 3.0T data only.

Results

The ADC values measured at the 1.5T system and the temperature-corrected ADC values from the 3.0T system are presented in Supplementary Table 1 (online only) for the two ROI selections. Apart from zoomed EPI and readout-segmented EPI, the differences of sample-specific ADC values measured with 1200 mm^2 ROIs were $<0.05 \times 10^{-3} \text{ mm}^2/\text{s}$ compared to the conventional series at the same field strength. The values at 1.5T were slightly higher ($0.05 \times 10^{-3} \text{ mm}^2/\text{s}$ on the average) than at 3.0T. The readout-segmented and zoomed EPI sequences produced slightly higher ADC values ($0.07 \times 10^{-3} \text{ mm}^2/\text{s}$ and $0.03 \times 10^{-3} \text{ mm}^2/\text{s}$ on the average, respectively) than the conventional sequences. The standard deviations of the 1200 mm^2 ROI pixel values were $\leq 0.12 \times 10^{-3} \text{ mm}^2/\text{s}$ in all measurements. With 3800 mm^2 ROIs, the highest SD values were $0.16 \times 10^{-3} \text{ mm}^2/\text{s}$ for 1.5T and $0.51 \times 10^{-3} \text{ mm}^2/\text{s}$ for 3.0T. The highest SD values originated from B_1 artifacts. Variation of approximately $0.1 \times 10^{-3} \text{ mm}^2/\text{s}$ in ADC values as a function of slice number was observed, as well as sequence-dependent differences in this measurement (Figs. 3 and 4).

The SNRs of the trace-weighted images ($b = 800 \text{ s/mm}^2$) are presented in Table 3 for each technique and field strength. The SNR variability was larger at 3.0T than at 1.5T, partly affected by increased artifacts. The use of coil intensity normalization caused notable changes in relative signal levels between different samples, i.e. the contrast was changed when this option was used.

In the visual assessment of the artifacts, a relative artifact level index was calculated for each acquisition technique (conventional EPI = 100). The lowest index, meaning the lowest level of artifacts, was calculated for readout-segmented EPI (=64), followed by zoomed EPI (=78) and patient-specific B_1 shim (=89). The images acquired without parallel imaging (=118) or with coil intensity normalization (=118) were given higher indices than conventional EPI. The positive effects of the patient-specific B_1 shim function were clear at 3.0T (Fig. 5). The geometric distortion, blurring and N/2 artifacts were mostly absent from the readout-segmented EPI images (Fig. 6). Zoomed EPI produced mostly artifact-free images, apart from B_1 artifact similar to conventional and readout-segmented EPI. An example of edge detection image is presented in Fig. 7, demonstrating stronger artifacts in a 3.0T image compared with a 1.5T image.

Discussion

The aim of this study was to construct a body-sized phantom for the evaluation of diffusion-weighted body MR image quality and ADC values with different

Fig. 3. Slice by slice measurement of sample ADC values in 1.5T from two consecutive acquisitions. Values of the first and last slices may have been affected by the edge of the sample.

Fig. 4. Slice-by-slice measurement of sample ADC values in 3.0T from two consecutive acquisitions 1 and 2. Values of the first and last slices may have been affected by the edge of the sample. The zoomed EPI series had a restricted coverage in slice direction and the slice numbers are shifted to match the actual location of the phantom with other measurements.

acquisition options at 1.5T and 3.0T. The agreement of ADC values was generally good, but differences between conventional and readout-segmented EPI acquisition as well as off-center variation in slice direction were observed. The phantom demonstrated well the lower level of image artifacts in acquisitions using patient-specific B_1 shim, readout-segmented EPI, or zoomed EPI.

Although a body-sized phantom can be considered impractical (6), the phantom construction was a straight-forward procedure. The size of the phantom

was essential to reflect abdominal imaging conditions and especially in demonstrating the artifacts related to the inhomogeneous B_1 field. The observed ADC values differed slightly from the expected. This may have been caused by slight modifications to the manufacturing process of the gels compared to Lavdas et al. (5).

We observed a small difference in the ADC values between 1.5T and 3.0T systems, which is in concordance with the values reported by Lavdas et al. for the same vendor (6). In our study the temperature differences, although corrected, may have caused uncertainty

Table 3. Signal-to-noise ratios (SNR) of diffusion-weighted images ($b = 800 \text{ s/mm}^2$) of individual samples 1–4 for different acquisition options.

Field strength and image	Acquisition option	SNR of individual samples			
		1	2	3	4
1.5T, $b = 800 \text{ s/mm}^2$	Conventional series	34	36	32	16
	–without parallel imaging	21	35	30	17
	–with coil intensity normalization	28	41	32	16
	Readout-segmented EPI	30	44	42	17
3.0T, $b = 800 \text{ s/mm}^2$	Conventional series	46	87	100	62
	–without parallel imaging	26	82	100	62
	–with coil intensity normalization	50	101	107	88
	–with patient-specific B_1 shim	50	70	73	53
	Readout-segmented EPI	59	137	121	69
	Zoomed EPI	33	60	63	46

EPI, echo planar imaging.

Fig. 5. The effect of patient specific B_1 shim option (right) in trace-weighted image ($b = 400 \text{ s/mm}^2$) at 3.0T.

in the observed ADC values. The difference between the room temperatures at the beginning and at the end of the measurements was $\leq 0.5^\circ\text{C}$. However, the phantom may have warmed up in the system bore during the measurements, which could be detected only by measuring the temperature directly from the liquid inside the phantom. Also, the ADC temperature correction may be not accurately valid for the gels; however, the temperature differences in our study were small. We believe that the accuracy of controlling the phantom temperature was sufficient for this study.

A suitable phantom could be used in calibration of the ADC values in an imaging center using several MRI systems, which in turn would facilitate the use of the values, e.g. in the therapy response evaluation. There are, however, intra-system problems as well. The slice-by-slice variation of the ADC was in line with the observation of Malyarenko et al., who have reported the measurements being less reproducible

across scanners in the off-center region, explained by gradient non-linearities (4). This issue affects especially the DWI of liver, because a large part of the liver can be located near the edge of the FOV. In our study gradient non-linearity could explain the trends observed in Figs. 3 and 4, but not the differences between conventional and readout-segmented EPI sequences. These differences require further investigation. Significant differences in the ADC values between conventional and readout-segmented EPI were not observed in other recent phantom or patient studies (20,21).

The efforts for the standardization of the DWI have focused on the ADC, even if reviewing of the trace-weighted images is at least equally important in clinical routine. We observed variability in SNR and contrast in the trace-weighted images acquired with different imaging options. Although coil intensity normalization had only minor effects on the ADC values, it altered the contrast of individual objects substantially in the

Fig. 6. A trace-weighted image ($b = 0 \text{ s/mm}^2$) and an ADC map acquired with readout-segmented EPI sequence at 1.5T.

Fig. 7. Edge detection images reflecting image intensity gradient magnitude, calculated from trace-weighted images ($b = 800 \text{ s/mm}^2$) including coil intensity normalization option at 1.5T (left) and 3.0T (right). With this option there are no strong intensity gradients near the coil elements on the phantom surface area. The increased intensity in the central area of the image arises from increased noise due to signal (and noise) amplification, and partly from parallel imaging. The intensity gradients on the sample areas are stronger at 3.0T. The chemical shift effects and distortion are also larger at 3.0T, as expected. Note the effect of a distortion correction filter seen on the edges of the images.

trace-weighted images. The use of coil intensity normalization should therefore be standardized in clinical protocols.

The image artifacts arose from the basic properties of the EPI sequence (N/2-type artifacts, blurring, and geometric distortion), the use of parallel imaging, and B_1 inhomogeneity at 3.0T. The use of patient-specific B_1 shim function enabled reduction of the artifacts. The use of parallel imaging is beneficial in abdominal DWI, but the artifacts can be problematic in the ADC measurements. These artifacts are probably pronounced in phantom images. The effects of parallel imaging on ADC measurements have not been reported in human studies (22,23). The typical artifacts of the conventional EPI sequence were largely removed by using the readout-segmented EPI sequence in this study. Based on these encouraging results, the clinical feasibility of this sequence in body area should be closer studied, despite the longer imaging time and possible problems

with motion. Also zoomed EPI has clinical potential in artifact reduction, shown also in clinical studies published so far (24–27).

The number of ADC measurements was limited, thus the results of the ADC differences were preliminary in nature. The measures to control the effects of temperature dependency were also limited, which could be addressed by replacing one of the samples with a bottle of pure water to act as a standard of reference. Intra-system repeatability of the phantom ADC results was not measured. However, Lavdas et al. have found the ADC intra-system repeatability to be excellent for similar sample compositions (5). Finally, our phantom study offered a limited view on clinical DWI. The perfusion effect or motion were not present in the phantom, and large volume of liquid and sharp interfaces may have produced stronger artifacts than a real patient.

In conclusion, the evaluation of diffusion-weighted MRI data with the body-sized phantom showed good

agreement of ADC values between different acquisition options and field strengths. The use of patient-specific B₁ shim, readout-segmented EPI, or zoomed EPI reduced the artifacts of DWI in this study.

Acknowledgements

The authors would like to thank Matti Keinänen, laboratory technician, for his help in constructing the phantom.

Declaration of conflicting interests

The author(s) declared no potential conflicts of interest with respect to the research, authorship, and/or publication of this article.

Funding

The author(s) received no financial support for the research, authorship, and/or publication of this article.

References

1. Padhani AR, Liu G, Mu-Koh D, et al. Diffusion-weighted magnetic resonance imaging as a cancer biomarker: consensus and recommendations. *Neoplasia* 2009;11:102–125.
2. Delakis I, Moore EM, Leach MO, et al. Developing a quality control protocol for diffusion imaging on a clinical MRI system. *Phys Med Biol* 2004;49:1409–1422.
3. Chenevert TL, Galbán CJ, Ivancevic MK, et al. Diffusion coefficient measurement using a temperature-controlled fluid for quality control in multicenter studies. *J Magn Reson Imaging* 2011;34:983–987.
4. Malyarenko D, Galbán CJ, Londy FJ, et al. Multi-system repeatability and reproducibility of apparent diffusion coefficient measurement using an ice-water phantom. *J Magn Reson Imaging* 2013;37:1238–1246.
5. Lavdas I, Behan KC, Papadaki A, et al. A phantom for diffusion-weighted MRI (DW-MRI). *J Magn Reson Imaging* 2013;38:173–179.
6. Lavdas I, Miquel ME, McRobbie DW, et al. Comparison between diffusion-weighted MRI (DW-MRI) at 1.5 and 3 Tesla: a phantom study. *J Magn Reson Imaging* 2014;40:682–690.
7. Tofts PS, Lloyd D, Clark CA, et al. Test liquids for quantitative MRI measurements of self-diffusion coefficient in vivo. *Magn Reson Med* 2000;43:368–374.
8. Laubach HJ, Jakob PM, Loeblad KO, et al. A phantom for diffusion-weighted imaging of acute stroke. *J Magn Reson Imaging* 1998;8:1349–1354.
9. Matsuya R, Kuroda M, Matsumoto Y, et al. A new phantom using polyethylene glycol as an apparent diffusion coefficient standard for MR imaging. *Int J Oncol* 2009;35:893–900.
10. Miquel ME, Scott AD, Macdougall MD, et al. In vitro and in vivo repeatability of abdominal diffusion-weighted MRI. *Br J Radiol* 2012;85:1507–1512.
11. Kangarlu A, Baertlein BA, Lee R, et al. Dielectric resonance phenomena in ultra high field MRI. *J Comput Assist Tomogr* 1999;23:821–831.
12. Dietrich O, Reiser MF, Schoenberg SO. Artifacts in 3-T MRI: Physical background and reduction strategies. *Eur J Radiol* 2008;65:29–35.
13. Katscher U, Börner P. Parallel RF transmission in MRI. *NMR Biomed* 2006;19:393–400.
14. Pfeuffer J, van de Moortele PF, Yacoub E, et al. Zoomed functional imaging in the human brain at 7 Tesla with simultaneous high spatial and high temporal resolution. *NeuroImage* 2002;17:272–286.
15. Rieseberg S, Frahm J, Finsterbusch J. Two-dimensional spatially-selective RF excitation pulses in echo-planar imaging. *Magn Reson Med* 2002;47:1186–1193.
16. Porter DA, Heidemann RM. High resolution diffusion-weighted imaging using readout-segmented echo planar imaging, parallel imaging and a two-dimensional navigator-based reacquisition. *Magn Reson Med* 2009;62:468–475.
17. Dietrich O, Raya JG, Reeder SB, et al. Measurement of signal-to-noise ratios in MR images: influence of multi-channel coils, parallel imaging, and reconstruction filters. *J Magn Reson Imaging* 2007;26:375–385.
18. Schneider CA, Rasband WS, Eliceiri KW. NIH Image to ImageJ: 25 years of image analysis. *Nature Methods* 2012;9:671–675.
19. Le Bihan D, Delannoy J, Levin RL. Temperature mapping with MR of molecular diffusion: application to hyperthermia. *Radiology* 1989;171:853–857.
20. Koyasu S, Iima M, Umeoka S, et al. The clinical utility of reduced-distortion readout-segmented echo-planar imaging in the head and neck region: initial experience. *Eur Radiol* 2014;24:3088–3096.
21. Thian YL, Xie W, Porter DA, et al. Readout-segmented echo-planar imaging for diffusion-weighted imaging in the pelvis at 3T – a feasibility study. *Acad Radiol* 2014; 21:531–537.
22. Taouli B, Martin AJ, Qayyum A, et al. Parallel imaging and diffusion tensor imaging for diffusion-weighted MRI of the liver: preliminary experience in healthy volunteers. *Am J Roentgenol* 2004;183:677–680.
23. Yoshikawa T, Kawamitsu H, Mitchell DG, et al. ADC measurement of abdominal organs and lesions using parallel imaging technique. *Am J Roentgenol* 2006;187: 1521–1530.
24. Riffel P, Michaely HJ, Morelli JN, et al. Zoomed EPI-DWI of the head and neck with two-dimensional, spatially-selective radiofrequency excitation pulses. *Eur Radiol* 2014;24:2507–2512.
25. Seeger A, Klose U, Bischof F, et al. Zoomed EPI DWI of acute spinal ischemia using a parallel transmission system. *Clin Neuroradiol* 2014; Epub ahead of print.
26. Thierfelder KM, Sommer WH, Dietrich O, et al. Parallel-transmit-accelerated spatially-selective excitation mri for reduced-fov diffusion-weighted-imaging of the pancreas. *Eur J Radiol* 2014;83:1709–1714.
27. Thierfelder KM, Scherr MK, Notohamiprodjo M, et al. Diffusion-weighted MRI of the prostate: advantages of zoomed EPI with parallel-transmit-accelerated 2D-selective excitation imaging. *Eur Radiol* 2014;24: 3233–3241.