<https://helda.helsinki.fi>

Parameters governing the performance of electrical mobility spectrometers for measuring sub-3 nm particles

Cai, Runlong

2019-01

Cai , R , Jiang , J , Mirme , S & Kangasluoma , J 2019 , ' Parameters governing the performance of electrical mobility spectrometers for measuring sub-3 nm particles ' , Journal of Aerosol Science , vol. 127 , pp. 102-115 . <https://doi.org/10.1016/j.jaerosci.2018.11.002>

<http://hdl.handle.net/10138/321496>

<https://doi.org/10.1016/j.jaerosci.2018.11.002>

cc_by_nc_nd

acceptedVersion

Downloaded from Helda, University of Helsinki institutional repository.

This is an electronic reprint of the original article.

This reprint may differ from the original in pagination and typographic detail.

Please cite the original version.

**Parameters governing the performance of electrical mobility spectrometers for measuring
sub-3 nm particles**

Runlong Cai^{a,b*}, Jingkun Jiang^a, Sander Mirme^c, and Juha Kangasluoma^{b,d}

^a State Key Joint Laboratory of Environment Simulation and Pollution Control, School of
Environment, Tsinghua University, 100084 Beijing, China;

^b Institute for Atmospheric and Earth System Research / Physics Faculty of Science, University of
Helsinki, P.O. Box 64, 00014 Helsinki, Finland

^c Institute of Physics, Laboratory of Environmental Physics, University of Tartu, Ülikooli 18, EE-
50090 Tartu, Estonia

^d Aerosol and Haze Laboratory, Beijing Advanced Innovation Center for Soft Matter Science and
Engineering, Beijing University of Chemical Technology, 100029 Beijing, China

Email of the corresponding author: crl14@mails.tsinghua.edu.cn

To be submitted to
Journal of Aerosol Science

October 25, 2018

26 **Abstract**

27 Measuring aerosol size distributions accurately down to ~1 nm is a key to nucleation studies, and it
28 requires developments and improvements in instruments such as electrical mobility spectrometers in
29 use today. The key factors characterizing the performance of an electrical mobility spectrometer for
30 sub-3 nm particles are discussed in this study. A parameter named as Π is proposed as a figure of
31 merit for the performance of an electrical mobility spectrometer in the sub-3 nm size range instead of
32 the overall detection efficiency. Π includes the overall detection efficiency, the measurement time in
33 each size bin, the aerosol flow rate passing through the detector, and the aerosol-to-sheath flow ratio
34 of the differential mobility analyzer. The particle raw count number recorded by the detector can be
35 estimated using Π at a given aerosol size distribution function, $dN/d\log d_p$. The limit of detection for
36 the spectrometer and the statistical uncertainty of the measured aerosol size distribution can also be
37 readily estimated using Π . In addition to Π , the size resolution of an electrical mobility analyzer is
38 another factor characterizing the systematic errors originated from particle sizing. Four existing
39 electrical mobility spectrometers designed for measuring sub-3 nm aerosol size distributions,
40 including three scanning/differential mobility particle spectrometers and one differential mobility
41 analyzer train, are examined. Their optimal performance is evaluated using Π and the size resolution.
42 For example, the Π value and the size resolution of a diethylene-glycol differential mobility particle
43 spectrometer for 1.5 nm particles are $8.0 \times 10^{-4} \text{ cm}^3$ and 5.7, respectively. The corresponding relative
44 uncertainty of the measured size distribution is approximately 9.6% during an atmospheric new
45 particle formation event with a $dN/d\log d_p$ of $5 \times 10^5 \text{ cm}^{-3}$. Assuming an adjustable sheath flow rate of
46 the differential mobility analyzer, the optimal size resolution is approximately 5-9 when measuring
47 atmospheric new particle formation events.

48 **Keywords:** electrical mobility spectrometer; sub-3 nm particles; penetration efficiency; resolution;
49 uncertainty; aerosol size distribution

50 **1. Introduction**

51 Measuring aerosol size distributions down to the cluster size (1.5 ± 0.4 nm) lays the foundation for
52 understanding the mechanisms of atmospheric new particle formation (Almeida et al. 2013; Chen et
53 al. 2012; Kulmala et al. 2012). Gaseous precursors in atmospheric environments form molecular
54 clusters via random collisions, and the clusters subsequently grow via condensational growth into
55 larger particles if they are sufficiently stable (Zhang et al. 2012) and are not scavenged by larger
56 particles due to coagulation (Kerminen, Pirjola and Kulmala 2001; McMurry 1983). Measuring
57 particles accurately down to the cluster size has been a challenging task, yet recent progresses in their
58 measurements have provided insights into atmospheric new particle formation (e.g., Kirkby et al.
59 2016; Sipilä et al. 2016). In addition, sub-3 nm processes in industrial applications have been recently
60 studied with new measurements (Alanen et al. 2015; Nosko, Vanhanen and Olofsson 2016; Tang et
61 al. 2017).

62 Developments of sub-3 nm particle measurement techniques in the past decade have made it feasible
63 to observe aerosol size distributions down to cluster sizes at ambient concentrations. Iida, Stolzenburg
64 and McMurry (2009) proposed to use diethylene glycol (DEG) as the particle counter working fluid,
65 which helped to extend the lower size limit of the ultrafine condensation particle counter (UCPC,
66 Stolzenburg and McMurry 1991) and the particle size magnifier (PSM, Kim, Okuyama and Fernández
67 de la Mora 2003; Kogan and Burnasheva 1960; Okuyama, Kousaka and Motouchi 1984; Vanhanen
68 et al. 2011) down to ~ 1 nm. Using a DEG-UCPC as the particle detector, a DEG-scanning mobility
69 size spectrometer (DEG-SMPS, Jiang et al. 2011b) can measure sub-3 nm aerosol size distributions
70 during atmospheric new particle formation events (Cai et al. 2017b; Jiang et al. 2011c). Compared to
71 instruments using other techniques to measure sub-3 nm aerosol size distributions, e.g., the pulse-
72 height CPC (Kuang 2018; Sipilä et al. 2008; Sipilä et al. 2009; Weber et al. 1998), the scanning PSM
73 (Lehtipalo et al. 2014; Vanhanen et al. 2011), the neutral cluster and air ion spectrometer (NAIS,
74 Manninen et al. 2009), and the symmetric inclined grid mobility analyzer (Tammet 2011), an

75 electrical mobility spectrometer using a CPC as the detector usually has a higher size resolution and
76 lower background noise.

77 However, the current electrical mobility spectrometers for sub-3 nm particles still need improvements
78 towards higher temporal resolutions, higher size resolutions, higher sensitivities with respect to
79 aerosol size distributions, and lower uncertainties in atmospheric conditions. The sampling efficiency,
80 aerosol charge fraction, the penetration efficiencies through the charger and the differential mobility
81 analyzer (DMA), and the detection efficiency of the CPC decrease sharply with decreasing particle
82 size in sub-2 nm size range (Jiang et al. 2011b). These low efficiencies limit the overall detection
83 efficiency of an electrical mobility spectrometer, leading to high limits of detection (LODs) and large
84 uncertainties at typical atmospheric particle number concentrations. As a result, the concentration of
85 the newly formed particles may not surpass the LOD of the instrument. For example, the DEG-SMPSs
86 recorded only several raw counts when measuring ~1.3 nm particles in two reported campaigns (Cai
87 et al. 2017b; Jiang et al. 2011c) during relatively strong new particle formation events ($dN/d\log d_p >$
88 10^5 cm^{-3}). The unknown chemical composition of ambient new particles, together with the assumed
89 charge fractions for sub-3 nm particles, also causes non-negligible uncertainties (Jiang et al. 2011b;
90 Kangasluoma and Kontkanen 2017; Kangasluoma et al. 2016b).

91 Several improvements in the electrical mobility spectrometers for measuring sub-3 nm particles have
92 been suggested recently. Sampling from the core of the flow minimizes the particle losses in the
93 sampling line (Kangasluoma et al. 2016a). A miniature cylindrical DMA operating at a medium
94 sheath flow rate (20-30 liters per minute) was developed to improve the penetration efficiency and
95 size resolution for sub-3 nm particles (Cai et al. 2017a). A supercritical DMA working at reduced
96 aerosol and sheath flow rates (Cai et al. 2018a; Fernández de la Mora 2017) was recently applied to
97 improve the size resolution of a differential mobility particle spectrometer (Kangasluoma et al. 2018).

98 A new DMA (TSI Inc., model 3086) to classify particles down to ~1 nm was developed and used for
99 the commercialized DEG-SMPS system (Stolzenburg et al. 2018). A DMA-train composed of six

DMAs operating in parallel at fixed but different particle sizes was developed to improve the temporal resolution (Stolzenburg, Steiner and Winkler 2017).

However, the overall detection efficiency together with the size resolution as reported in previous studies (e.g., Jiang et al. 2011b; Kangasluoma et al. 2018; Stolzenburg, Steiner and Winkler 2017) cannot fully characterize the potential improvements to the instruments in the sub-3 nm size range. Increasing the raw count number recorded by the detector CPC will reduce the statistical uncertainty and the LOD. Therefore, increasing the measurement time in each size bin and the aerosol flow rate passing through the detector will also increase the raw count number and hence improve the instrument performance (Iida 2008). Thus, one should also take the measurement time and the aerosol flow rate passing through the detector into account when evaluating the performance of an electrical mobility spectrometer. Furthermore, a high size resolution is preferred to obtain more accurate aerosol size distributions, whereas the relationship between size resolution and the systematic errors of the inverted sub-3 nm aerosol size distribution originated from uncertainties in particle sizing has not been quantitatively discussed before.

In this study, we propose a Π parameter, i.e., a product including the overall particle detection efficiency, the measurement time in each size bin, the aerosol flow rate passing through the detector, and the integral of DMA transfer function with respect to particle diameter, as a figure of merit for characterizing the performance of an electrical mobility spectrometer for sub-3 nm particles instead of the overall detection efficiency (P). Four existing electrical mobility spectrometers are theoretically compared to illustrate the advantage of Π over P . The influence of Π and the size resolution on the uncertainty of the inverted sub-3 nm aerosol size distribution is theoretically analyzed. The feasibility of using a Faraday cage electrometer (FCE) instead of a CPC as the detector of an electrical mobility spectrometer and the systematic error of a scanning PSM are also discussed based on analyses using Π and the size resolution.

2. Theory

2.1 Definition of Π

A typical electrical mobility spectrometer consists of a sampling line, a charger (neutralizer), a DMA to classify (select) particles with targeted electrical mobility, and a CPC as a detector (Wang and Flagan 1990). Assuming that the DMA sampling flow rate is equal to the aerosol flow rate exiting the DMA, the relationship between the expected value of CPC raw count number and the aerosol size distribution function is:

$$C^* = \int_{-\infty}^{+\infty} P(d_p) \cdot \Omega(d_p^*, d_p) \cdot n(\log d_p) \cdot t \cdot Q \cdot d\log d_p, \quad (1)$$

where C^* is the number of particles counted by the CPC at the DMA centroid electrical mobility diameter, d_p^* ; P is the overall detection efficiency of the electrical mobility spectrometer (see Eq. 2 below); Ω is the diffusional DMA transfer function (Stolzenburg 2018; Stolzenburg and McMurry 2008); $n(\log d_p)$ is the aerosol size distribution function (probability density function of particle number concentration) using the logarithm to the base 10, $dN/d\log d_p$; t is the measurement time or averaging time for each size bin; and Q is the effective aerosol flow rate. The effective aerosol flow rate is defined as the equivalent undiluted aerosol flow rate passing through the detector. For example, the effective aerosol flow rate of the DEG-UCPC in Jiang et al. (2011b) is the capillary flow rate (0.03 liter per minute, lpm) rather than the inlet flow rate (2 lpm) or the total flow rate passing through the optical detector (0.3 lpm).

In previous studies (e.g., Jiang et al. 2011b; Stolzenburg, Steiner and Winkler 2017), P is usually determined empirically as the product of the efficiencies for each part of the instrument, i.e.,

$$P = \eta_{\text{samp}} \cdot \eta_{\text{char}} \cdot f_c \cdot \eta_{\text{DMA}} \cdot \eta_{\text{cpc}} \cdot \eta_{\text{corr}}, \quad (2)$$

where η_{samp} , η_{char} , and η_{DMA} are the penetration efficiencies passing through the sampling line, the charger, and the DMA, respectively; f_c is the charge fraction for singly charged particles (the multi-charging effect for sub-3 nm particles can be neglected); η_{cpc} is the detection efficiency of the CPC; and η_{corr} is a correction factor characterizing additional particle losses due to flow bending (Wang, Flagan and Seinfeld 2002), flow splitting (e.g., 0.8 in Kangasluoma et al. (2018)), and the difference

150 between the calibrated and true efficiencies because of the different particle spatial distributions in
 151 the flow field for the calibration setup and the electrical mobility spectrometer.

152 The performance of an electrical mobility spectrometer for sub-3 nm particles was previously
 153 characterized using P . However, as shown in Eq. 1, one has to know t , Q , and Ω in addition to P alone
 154 to estimate C^* . In this study, we propose a Π parameter as a product of P , t , Q , and parameters of the
 155 DMA transfer function for better characterizing the performance of an electrical mobility
 156 spectrometer. Π is defined as:

$$157 \quad \Pi = P \cdot t \cdot Q \cdot \beta \cdot \left(-\frac{d \log d_p}{d \log Z_p} \right), \quad (3)$$

158 where β is the aerosol-to-sheath flow ratio of the DMA; Z_p is the corresponding electrical mobility of
 159 particles with the diameter d_p and a certain charge. Although $-d \log d_p / d \log Z_p$ is only determined by
 160 particle diameter when assuming all the charged sub-3 nm particles can only be singly charged, we
 161 include it in the definition of Π for a more straightforward relationship between the aerosol size
 162 distribution function and the CPC raw count number. According to Eqs. 1 and 3, the relationship
 163 between C^* and $n(\log d_p)$ can be expressed as:

$$164 \quad C^* = \int_0^{+\infty} P(Z_p) \cdot \Omega(Z_p^*, Z_p) \cdot n(\log d_p) \cdot t \cdot Q \cdot \left(-\frac{d \log d_p}{d \log Z_p} \right) \cdot \frac{1}{\ln 10} \cdot \frac{1}{Z_p} \cdot dZ_p, \quad (4)$$

$$= \frac{\Pi^*}{\ln 10 \cdot \beta} \cdot \int_0^{+\infty} \frac{P(Z_p)}{P(Z_p^*)} \cdot n(\log d_p) \cdot \frac{1}{Z_p} \cdot \Omega(Z_p^*, Z_p) \cdot dZ_p$$

165 where Π^* and Z_p^* are the Π parameter and the electrical mobility at d_p^* , respectively, and -
 166 $d \log d_p / d \log Z_p$ is approximated as a constant value in the mobility window of Ω where Ω is nonzero
 167 (or more specifically, where Ω is larger than a given small value because Ω is always nonzero by its
 168 expression). Considering that $\int_0^{+\infty} \Omega(Z_p^*, Z_p) \cdot dZ_p = \beta \cdot Z_p^*$, one can find that β and $-d \log d_p / d \log Z_p$
 169 in Eq. 3 actually come from the integral of the diffusional transfer function with respect to $\ln d_p$, i.e.,

$$170 \quad \int_{-\infty}^{+\infty} \Omega(d_p^*, d_p) \cdot d \ln d_p \approx -\frac{d \log d_p}{d \log Z_p} \cdot \beta. \text{ Assuming } P(Z_p), n(\log d_p), \text{ and } Z_p \text{ are constant values in the}$$

mobility window of Ω , the relationship between C^* and $n(\log d_p)$ can be approximated as (Knutson 1976; Stolzenburg and McMurry 2008):

$$C^* = \Pi^* \cdot n(\log d_p^*) / \ln 10 = \Pi^* \cdot n(\ln d_p^*), \quad (5)$$

As shown in Eqs. 4 and 5, Π is essentially a parameter relating the aerosol size distribution function and the expected value of CPC raw count number. Note that if t remains unchanged, increasing the width of each size bin of an electrical mobility spectrometer cannot increase C^* because the fraction of aerosol size distribution function contributing to C^* is determined by the DMA transfer function width rather than the bin width.

2.2 Size resolution and DMA transfer function

The dimensionless diffusional DMA transfer function is shown in Eq. 6 (Stolzenburg 2018; Stolzenburg and McMurry 2008),

$$\Omega = \frac{\sigma_\Omega}{\sqrt{2}\beta(1-\delta)} \left[\varepsilon \left(\frac{(\overline{Z_p} - 1) + \beta}{\sqrt{2}\sigma_\Omega} \right) + \varepsilon \left(\frac{(\overline{Z_p} - 1) - \beta}{\sqrt{2}\sigma_\Omega} \right) - \varepsilon \left(\frac{(\overline{Z_p} - 1) + \beta\delta}{\sqrt{2}\sigma_\Omega} \right) - \varepsilon \left(\frac{(\overline{Z_p} - 1) - \beta\delta}{\sqrt{2}\sigma_\Omega} \right) \right] \quad (6)$$

where σ_Ω is the dimensionless diffusion parameter; δ is the normalized flow imbalance and it is assumed to be zero in this study; $\overline{Z_p}$ is the dimensionless electrical mobility, Z_p/Z_p^* ; and ε is the error function.

The DMA resolution, R , determines the size resolution of an electrical mobility spectrometer. R is defined as the ratio of the centroid electrical mobility to the full width at half-maximum height of the electrical mobility based transfer function (Flagan 1999), i.e.,

$$R = \frac{Z_p^*}{\Delta Z_{\text{FWHM}}}, \quad (7)$$

where ΔZ_{FWHM} is the full width at half-maximum height of the electrical mobility. One can predict σ_Ω and the diffusional transfer function using R at a given β . σ_Ω can be inferred using Eq. 8 when knowing the $\sigma_{\Omega 0}$ value at any specific electrical mobility, Z_{p0} (Stolzenburg 2018).

$$\sigma_{\Omega} = \sigma_{\Omega 0} \cdot \sqrt{Z_p / Z_{p0}} \quad (8)$$

2.3 Uncertainties

Many factors may cause uncertainties/errors in the reported aerosol size distributions, e.g., deviations of the aerosol charge fractions from those calculated using the modified Fuchs theory (Hoppel and Frick 1986) for sub-3 nm particles; randomness in CPC counting; influence of particle chemical composition on the CPC detection efficiency (Jiang et al. 2011b; Kangasluoma and Kontkanen 2017; Kuang et al. 2012); and data inversion methods (Wiedensohler et al. 2012). The cause of CPC measurement uncertainties can be grouped into three main aspects: the statistical uncertainty caused by randomly distributed particles in a given volume which can be predicted using Poisson distribution (Iida 2008; Kangasluoma and Kontkanen 2017; Kuang et al. 2012); the uncertainties linearly proportional to particle concentration, such as the concentration fluctuation due to unstable flow rates and inhomogeneous air mass, which are the dominant uncertainties when the particle raw count number is high; and the CPC background noise due to potential leakage and homogeneous nucleation of the working fluid. In this study, we mainly focus on the CPC statistical uncertainty because it limits the optimal performance of an electrical mobility spectrometer in the sub-3 nm size range. The Π parameter can also characterize the relationship between C^* and $n(\log d_p)$ for particles larger than 3 nm. However, when measuring large particles (e.g., > 25 nm), the contribution of statistical uncertainty to the total uncertainties may be negligible due to the correspondingly large Π , and improving Π does not necessarily increase the performance of an electrical mobility spectrometer significantly. The CPC background noise will be accounted for when discussing the LODs. The statistical uncertainty predicted using Poisson distribution is shown in Eq. 9, where C and σ are the expected value and the standard deviation of the CPC raw count number, respectively.

$$\sigma = \sqrt{C} \quad (9)$$

3. Methods

217 3.1 Electrical mobility spectrometers

218 Four existing electrical mobility spectrometers are theoretically compared, including the first
219 prototype DEG-SMPS (Jiang et al. 2011b), an improved version of the DEG-SMPS (Cai et al. 2017a),
220 a differential mobility particle spectrometer (DMPS) using a supercritical DMA working at reduced
221 flow rates (Kangasluoma et al. 2018), and a DMA-train using six DMAs at different fixed voltages
222 with six downstream detectors (Stolzenburg, Steiner and Winkler 2017). In this study, the four
223 instruments are referred as the DEG-SMPS 2011, the DEG-SMPS 2017, the half-mini DMPS, and
224 the DMA-train. The commercialized TSI DEG-SMPS is not included since the penetration efficiency
225 for its DMA has not been reported, yet it can be estimated that the I value of the TSI DEG-SMPS is
226 between the I values of the DEG-SMPS 2011 and the DEG-SMPS 2017.

227 The specifications of the four instruments are shown in Table 1. Compared to the DEG-SMPS 2011,
228 the DEG-SMPS 2017 improves the sampling efficiency by taking aerosols from the core of sampling
229 flow (Kangasluoma et al. 2016a), deploys a miniature cylindrical DMA (mini-cyDMA) which has
230 higher penetration efficiency and size resolution for sub-3 nm particles (Cai et al. 2017a) compared
231 to the TSI nanoDMA (Chen et al. 1998; Jiang et al. 2011a) in the DEG-SMPS 2011 and increases the
232 capillary flow rate of the DEG-UCPC. The half-mini DMPS also uses the core sampling technique to
233 minimize particle diffusion loss in the sampling line. A supercritical DMA named half-mini DMA
234 (Fernández de la Mora and Kozlowski 2013) with reduced aerosol and sheath flow rates for
235 atmospheric measurements is deployed in the half-mini DMPS. Compared to the DEG-UCPC, the
236 Airmodus PSM has a higher effective aerosol flow rate, Q . There are six DMAs working at different
237 but fixed voltages in the DMA train (Stolzenburg, Steiner and Winkler 2017), and we adopt the
238 configurations for the sub-2.5 nm channels in this study. Although using a PSM at a relatively high
239 aerosol flow rate for the initial particle growth, the Q value of the DMA-train is limited by the
240 capillary flow rate of the downstream UCPC. The DMA-train was developed to achieve a high
241 temporal resolution. It can also be used to increase the CPC raw count number by averaging the

signals during a relatively long time interval (i.e., several minutes). To compare the uncertainties with other instruments at the same temporal resolution, we assume that the averaging time of the DMA-train is 300 s in this study. The scanning periods of the DEG-SMPS 2011, the DEG-SMPS 2017, and the half-mini DMPS are all 300 s in this study, which is a typical scanning period for an SMPS/DMPS in atmospheric measurements considering the scanning speed and the concentration of atmospheric aerosols. The charging properties for sub-3 nm particles are not well understood, and the aerosol charge fractions are different when using radioactive chargers and X-ray chargers due to different ion mobility (Chen and Jiang 2018; Jiang et al. 2014; Lee et al. 2005), yet in this study we simply use a widely used approximation for bipolar charge fractions (Wiedensohler 1988) for all the chargers for the purpose of comparison.

3.2 Test size distribution

The aerosol size distribution function to examine the uncertainties at different Δ values and size resolutions is a smoothed distribution obtained from atmospheric measurements (Jiang et al. 2011c). The aerosol size distribution function, $n(\log d_p)$, increases with the decreasing particle diameter (as shown in Fig. 5). Kangasluoma and Kontkanen (2017) estimated the uncertainties in sub-3 nm particle concentration measurements using a similar size distribution. The shape of the test size distribution affects the uncertainties of the measured $n(\log d_p)$. However, since this study mainly focuses on instrumental parameters, we use this distribution function measured during a typical atmospheric new particle formation event alone to illustrate the influence of size resolution on the systematic error due to the uncertainties in particle sizing. The test results using three other distributions including a uniform distribution, a power law distribution, and a lognormal distribution with a peak diameter of 2 nm are shown in the supplementary materials.

3.3 Inversion method

There are various methods to account for the broadening effect of a DMA (Ramachandran and Cooper 2011), i.e., to deal with the Fredholm integral equation of the first kind:

$$C^* = \int_{-\infty}^{+\infty} \frac{\Pi(Z_p)}{\beta} \cdot n(\log d_p) \cdot \Omega(Z_p^*, Z_p) \cdot d \log Z_p + P_{\text{rnd}}(C^*), \quad (10)$$

where $P_{\text{rnd}}(C^*)$ is a random number that follows Poisson distribution with an expected value of C^* .

Note that Eqs. 1, 4, and 10 are essentially the same except for the random error.

The data inversion method to recover aerosol size distributions from CPC raw count number in this study is a one-to-one linear inversion method using Eq. 5, i.e., assuming that Π and $n(\log d_p)$ are nearly constant over the mobility window of a given DMA. This one-to-one inversion method was used for DEG-SMPS data inversion in previous studies (e.g., Cai et al. 2017a; Jiang et al. 2011b). A proper inversion method can reduce the uncertainties in the inverted aerosol size distribution. However, some feasible methods previously used in SMPS/DMPS data inversion (such as the inversion routines using the least square method) may magnify the random errors in CPC raw counts into relatively large uncertainties (Enting and Newsam 1990). To constrain random errors, some inversion methods use smoothing steps in the inversion routines (e.g., Markowski 1987) or fit predefined size distribution functions to the data. However, these inversion methods may lead to bias in the shape of the inverted distribution function. Kandlikar and Ramachandran (1999) concluded that there is no single algorithm universally superior to the others. We tested the Hagen & Alofs method (Hagen and Alofs 1983) and the Expectation-Maximization algorithm (Maher and Laird 1985). When using these two inversion method, the uncertainties are not significantly smaller than using the one-to-one inversion method (Figs. S1 and S2). Thus, we use the one-to-one inversion method in this study for characterizing the relationship between uncertainties and size resolution.

The effect of scan speed on the inverted aerosol size distributions is assumed to be negligible in this study. When operating at a rapid scan speed, an SMPS may broaden and skew the measured aerosol size distributions if the effect of scan speed is not fully considered during data inversion. The finite response time of a CPC (Russell, Flagan and Seinfeld 1995) and the deviation for the static transfer function at a constant voltage from the scanning transfer function (Mai and Flagan 2018) are the major reasons. The broadening and skewing effects are usually significant during rapid scans (e.g.,

scan time < 20 s, Tröstl et al. 2015). However, to reduce the LODs and uncertainties for sub-3 nm particles, the scan speed is usually limited to obtain a relatively large II . For example, if approximating the DEG-UCPC 95% response time with 1.94 s (Enroth et al. 2018), the θ_s values (see Russell, Flagan and Seinfeld 1995) for the DEG-SMPS 2011 and the DEG-SMPS 2017 are approximately 0.0075 and 0.0094, respectively, indicating that neglecting the effect of the scan speed in this study will not lead to significant errors.

3.4 Analyses

The analyses conducted in this study are summarized below:

- The II values of the four electrical mobility spectrometers, their particle raw count number at a constant $n(\log d_p)$ of $5 \times 10^5 \text{ cm}^{-3}$, and the corresponding relative standard deviations;
- The statistical uncertainty and limit of detection as a function of II ;
- The random errors at different II values when measuring the test distribution;
- The systematic errors at different size resolutions when measuring the test distribution;
- The total uncertainties at different II values and size resolutions;
- The performance of FCE-based electrical mobility spectrometers and a scanning PSM characterized using II and size resolution.

4. Results and discussion

4.1 II parameter

The II parameter characterizes the performance of an electrical mobility spectrometer in the sub-3 nm size range better than the overall detection efficiency, P . According to Eqs. 5 and 9, a high CPC raw count number, C^* , is desirable to reduce the statistical uncertainty, and C^* is linearly proportional to II . P is usually small for sub-2 nm particles due to low aerosol charge fractions, low CPC detection efficiencies, and low penetration efficiencies through the sampling line, the charger, and the DMA. However, a relatively small P can be compensated using long measurement time, t , and a large aerosol

316 flow rate, Q . As shown in Fig. 1, the value of P for the half-mini DMPS is lower than or almost the
317 same to the value of P for the DEG-SMPS 2011. However, the half-mini DMPS uses a PSM for the
318 initial particle growth, which increases Q to 19.0 times that of the laminar flow DEG-UCPC in the
319 DEG-SMPS 2011. Thus, the half-mini DMPS has a larger II value hence higher CPC raw counts and
320 smaller statistical uncertainties compared to the DEG-SMPS 2011 when measuring aerosols with the
321 same number concentration. Similarly, although the DMA-train is developed to improve the temporal
322 resolution, its relatively low P values in the sub-2 nm particle size range can be compensated if using
323 long averaging time.

324 The minimum measurement uncertainty, i.e., the statistical uncertainty, of an electrical mobility
325 spectrometer can be estimated using the II parameter. According to Eqs. 5 and 9, the statistical
326 uncertainty is linearly proportional to the square root of II (Fig. 2). For example, the statistical
327 uncertainty of the DEG-SMPS 2017 for 1.5 nm particles (with a II of $8.04 \times 10^{-4} \text{ cm}^3$) is 5.4% when
328 the test $n(\log d_p)$ is 10^6 cm^{-3} , while the uncertainty increases to 53.5% when the test $n(\log d_p)$ decreases
329 to 10^4 cm^{-3} .

330 The LODs can also be readily estimated using II . According to Eq. 5, the LOD of an SMPS/DMPS
331 is determined by II and the background noise of the CPC due to leakage, homogeneous nucleation of
332 the working fluid vapor, and other potential reasons. The background noise of a CPC is usually
333 reported in particle raw count rates or equivalent particle concentrations. For example, the
334 homogeneous nucleation rates of the DEG-SMPS 2011, DEG-SMPS 2017, and the half-mini DMPS
335 are smaller than 0.003, 0.003, and 0.067 s^{-1} , respectively. The reported background concentrations of
336 commercialized CPCs are typically 0.01 cm^{-3} or smaller. These reported background particle count
337 rates and concentrations are usually average values based on long-time measurements in stable
338 conditions, whereas the false count rates are perhaps larger in practical conditions due to unstable
339 particle sources and environments. Nevertheless, the background noise of a CPC usually does not
340 lead to averagely more than one count in each size bin of an SMPS/DMPS during a scanning cycle

(typically 300 s for totally 15-30 size bins). Because the CPC count number is always an integer and considering the demand for temporal resolution, it is reasonable to define the LOD of an SMPS/DMPS as the corresponding $dN/d\log d_p$ when the CPC records only one count (Iida 2008). Accordingly, the LOD of an SMPS/DMPS can be estimated using the formula $LOD = 1/I$ (Fig. 3). Because the statistical uncertainty and LOD of an SMPS/DMPS is a function of I , the optimal configuration of an electrical mobility spectrometer should be determined using I rather than P . For example, the relationship between the detection efficiency of 1.2 nm tungsten oxide particles and the inlet flow rate of a tuned butanol CPC for sub-3 nm particles (shown in Fig. 4) was discussed in Kangasluoma et al. (2015). The detection efficiency peak appeared at ~1.6 lpm, which was regarded as the optimal inlet flow rate. For ease of illustration, we assumed this CPC is used as the particle detector of the half-mini DMPS. The peak value of P appears at ~1.6 nm, however, I is still increasing with the increasing Q until 2.2 lpm, and the expected value of CPC raw count number at 2.2 lpm is eventually larger than at 1.6 lpm. Thus, the inlet flow rate for the CPC when implemented in the half-mini DMPS should be optimized at 2.2 lpm with a comparatively small P but the largest I , hence the lowest statistical uncertainty and LOD.

4.2 Testing the performance of an SMPS/DMPS using I and R

In this study, the performance of an electrical mobility spectrometer is characterized by the random error and the systematic error in the measured aerosol size distribution. The random error is the standard deviation of the measured aerosol size distribution and is estimated using the statistical uncertainty. The systematic error is the difference between the test size distribution and the expected value of the inverted size distribution. It originates from uncertainties in particle sizing due to finite size resolution of the electrical mobility spectrometer and potential bias of the inversion method. As discussed in the Method section, we use the one-to-one linear inversion method in this study because it will not magnify the random errors and it is the inversion method currently used in the DEG-SMPSs and the half-mini DMPS. For the test distribution, the expected value of the measured CPC raw count

number is obtained using Eq. 4 that accounts for the size resolution. The CPC raw number is used as the input of Eq. 5, i.e., the one-to-one linear inversion method, to obtain the inverted size distribution. The difference between the inverted size distribution and the test size distribution is taken as the systematic error. The systematic errors due to uncertainties in aerosol charge fractions and the particle composition dependence of CPC detection efficiencies are not considered here.

The total uncertainty of the measured size distribution is characterized using the sum of the absolute value of the systematic error and the random error. The random error and the systematic error are assumed to be independent of each other, i.e., it is assumed that size resolution does not affect the CPC raw count number significantly when estimating the random error. The validity of this assumption was examined for the test distribution and the maximum relative uncertainty of the estimated random error is less than 18% (when $R = 1$).

The II value mainly affects the random error. Increasing II will reduce the random error when measuring the same aerosol size distribution. For the test distribution, the relative random error of the DEG-SMPS 2011 for 1.3 nm particles is 30.5%. When using the DEG-SMPS 2017 which has an improved II value, the relative random error decreases to 15.5%. If one can further improve II to 10 times that of the DEG-SMPS 2017, e.g., by improving the aerosol charge fractions, the relative random error can be reduced to 5.0% (Fig. 5).

Improving the size resolution reduces the systematic error. The relative systematic errors at different diameters generally decrease with the increasing size resolution (Fig. 6). For example, the estimated relative systematic errors using the test distribution are 7.5%, 3.5%, and 2.0% when the resolution for 1.5 nm particles is 3, 5, and 8, respectively. Accordingly, a low size resolution may lead to non-negligible uncertainties, while higher size resolution (e.g., 7.9 for 1.5 nm particles if using the half-mini DMPS with reduced flow rates) reduces the systematic errors.

390 The aerosol size distribution function also affects the systematic errors; however, the influence is
 391 difficult to quantify using simple parameters. The relationships between the relative systematic errors
 392 of the inverted size distributions for 1.5 nm particles and the size resolution when measuring different
 393 aerosol size distributions are shown in Fig. S1. All the relative systematic errors decrease with the
 394 increasing size resolution.

395 Improving II and the size resolution, R , will reduce the uncertainties of the measured aerosol size
 396 distribution, as shown in Fig. 7a. The uncertainty is characterized using the sum of the random error
 397 (Fig. 5) and the absolute value of the systematic error (Fig. 6). Compared to the DEG-SMPS 2011,
 398 the other three instruments improve the optimum performances by increasing II and R . For example,
 399 the uncertainties of the inverted size distribution at 1.5 nm are 22.5%, 9.4%, 8.5%, and 10.5% when
 400 using the DEG-SMPS 2011, the DEG-SMPS 2017, the half-mini DMPS, and the DMA-train,
 401 respectively.

402 The II parameter and R can also help to determine the DMA sheath flow configuration. The value of
 403 II increases with an increasing DMA aerosol-to-sheath flow ratio (β); however, the size resolution
 404 decreases simultaneously. When varying the sheath flow rate in a reasonable range so that the DMA
 405 transfer function is not broadened due to turbulence, one can estimate the uncertainties at different
 406 flow configurations and accordingly figure out a an ideal sheath flow rate with a comparatively small
 407 uncertainty (see Fig. 7). Decreasing the DMA resolution below approximately 5 to obtain more
 408 particle raw counts may increase the total uncertainties at relatively high particle number
 409 concentrations due to the increasing systematic error. Conversely, increasing the DMA resolution
 410 above approximately 8-9 does not necessarily reduce the systematic errors effectively, while it still
 411 lowers the II value and increases the total uncertainties at relatively low particle number
 412 concentrations. The optimal size resolution is also affected by the shape of the aerosol size
 413 distribution, yet this effect is similar to the effect of particle number concentration. A size distribution
 414 corresponding to larger systematic errors and a higher aerosol concentration that causes smaller

random errors both decrease the optimal size resolution. For example, the optimal resolutions corresponding to the minimum uncertainties for the DEG-SMPS 2017, the half-mini DMPS, and the DMA train are 8.6, 8.6, and 8.3 when measuring a lognormal aerosol size distribution, respectively (Fig. S2b), similar to the optimal resolutions for the test distribution but at a higher $dN/d\log d_p$ of $3 \times 10^6 \text{ cm}^{-3}$ (Fig. S2c). Thus, we can conclude that the optimal DMA size resolution appears to be ~5-9 for 1.5 nm particles when measuring typical atmospheric new particle formation events and that the improvements to the current electrical mobility spectrometers should come from the improvements in Π , that is, mainly higher effective aerosol flow rates of the CPCs, higher DMA penetration efficiencies using specially designed DMAs, or improved aerosol charge fractions.

4.3 FCE-based electrical mobility spectrometers and the scanning PSM

The FCE-based electrical mobility spectrometers (e.g., the NAIS) and the scanning PSM that measures total particle concentrations at different supersaturation ratios are also used for sub-3 nm particle measurements. Although their working principles are different from that of a CPC-based electrical mobility spectrometer, the governing factors for their performances can also be analyzed using the Π parameter and the size resolution.

The advantage of an FCE over a CPC is that the detection efficiency of the FCE is unaffected by the chemical compositions of the measured particles, while the drawback of the FCE is its comparatively high background noise level. Different from a CPC, the LOD of an FCE is determined as 3 times the root mean square (RMS) value of the noise level based on the average over the measurement time. Zero drift is neglected since we assume that it can be accounted for by tuning the zero offset at the optimal performance. The relationship between the LOD of an FCE-based electrical mobility spectrometer and the Π parameter using the current definition is:

$$\text{LOD}(dN/d\log d_p) = \frac{\ln 10 \cdot 3 \cdot \text{RMS}_1 \cdot \sqrt{t \cdot 1 \text{ s}}}{e \cdot \Pi}, \quad (11)$$

438 where RMS_1 is the root mean square of the noise level in electric current based on a 1 s average, t is
 439 averaging time, and e is the elementary charge. Perhaps replacing t with $\sqrt{t}$ in the definition of I is
 440 more reasonable for an FCE-based electrical mobility spectrometer, yet we used the definition in Eq.
 441 3 for consistency. A new I parameter redefined for an FCE-based electrical mobility spectrometer
 442 and its relationship with the statistical error and LOD is discussed in the supplementary materials.
 443 Assuming that the CPCs of the four analyzed instruments are replaced with FCEs, the obtained LODs
 444 are not low enough to measure atmospheric new particle formation. The aerosol inlet flow rates of
 445 the simulated FCEs are assumed the same to the corresponding aerosol flow rate exiting the DMAs
 446 so that there is no dilution or bypass flow to reduce Q . The root mean square (RMS) of the noise level
 447 is assumed to be 0.1 fA based on 1 s average (Fernández de la Mora et al. 2017), which is a relatively
 448 low value among the present commercialized FCEs. The LOD of an imaginary SMPS is also
 449 estimated for evaluating the feasibility of measuring sub-3 nm particles using an FCE-based SMPS.
 450 We assume P and β are the same to those of the half-mini DMPS. The t and Q values of the imaginary
 451 SMPS are assumed to be 300 s and 25 lpm, respectively. As shown in Fig. 3b, the four instruments
 452 equipped with FCE as detector have too high LODs to measure atmospheric new particle formation
 453 events, (e.g., $n(\log d_p)$ was smaller than $\sim 10^6 \text{ cm}^{-3}$ in Jiang et al. (2011c)), since the LODs are
 454 magnitudes higher than the previous configurations using CPCs to detect particles. In addition, the
 455 LODs in the sub-1.5 nm size range in atmospheric measurements may be higher than estimated
 456 because of the influence of ions generated in the charger (Kallinger, Steiner and Szymanski 2012).
 457 Due to the large values of t and Q , the LOD of the imaginary SMPS for sub-2 nm particles approaches
 458 typical particle concentration during atmospheric new particle formation events. Assuming a
 459 background noise level of 0.01 fA (based on a 4 min average), the LOD of an NAIS can be estimated
 460 and is shown in Fig. 3b. The LOD of the NAIS is lower than the FCE-based SMPSs because it uses
 461 unipolar corona chargers instead of bipolar chargers and has a higher aerosol flow rate (27 lpm).

462 The relatively low size resolution of the scanning PSM limits its accuracy in measuring sub-3 nm
463 particles. Different from an electrical mobility spectrometer, the II value of a scanning PSM is usually
464 sufficiently high because of the high overall detection efficiency (P) due to no need for aerosol
465 charging. The measurement uncertainties of the scanning PSM is mainly determined by instrumental
466 and atmospheric instabilities rather than the statistical uncertainties. Accordingly, II is sufficiently
467 high for the scanning PSM so that improving II may not reduce the uncertainty. The size resolution,
468 however, is one of the limiting factors for the performance of the scanning PSM. The typical size
469 resolution for a scanning PSM (~1-4, defined differently from the size resolution of an electrical
470 mobility spectrometer) is relatively low (Cai et al. 2018b). As a result, the systematic error of a
471 scanning PSM is larger than typical electrical mobility spectrometers (see Fig. 6c). A proper inversion
472 method may help to reduce the systematic error of a scanning PSM (Cai et al. 2018b), yet the
473 performances of different inversion methods when inverting PSM data from atmospheric
474 measurements need to be further addressed.

475 5. Conclusions

476 A II parameter including the overall detection efficiency, measurement time, the effective aerosol
477 flow rate, and the aerosol-to-sheath flow ratio of the differential mobility analyzer, is proposed as a
478 figure of merit of an electrical mobility spectrometer for sub-3 nm particles, instead of the overall
479 detection efficiency alone. The statistical uncertainty and the limit of detection of an electrical
480 mobility spectrometer can be estimated using the II parameter. Together with the size resolution, II
481 can characterize the performance of an electrical mobility spectrometer for measuring sub-3 nm
482 particles. A larger II value and a higher size resolution reduce the total uncertainty of the measured
483 aerosol size distribution. Compared to a prototype DEG-SMPS, the other three instruments (the
484 modified DEG-SMPS, the half-mini DMPS, and the DMA-train) improved the II value and the size
485 resolution using different approaches. With the current SMPS/DMPS systems, the optimal DMA size
486 resolution was found to be ~5-9 when measuring typical atmospheric new particle formation events.

487 The Λ parameter and the size resolution also help to understand the performance of FCE-based
488 electrical mobility electrometers and scanning PSMs. To overcome the particle composition related
489 activation issue of CPCs, simply replacing the CPC in an existing SMPS/DMPS with an FCE will
490 not provide a sufficiently low detection limit for sub-3 nm particles in typical atmospheric conditions.
491 Thus, one has to improve the instruments towards higher Λ values if using an FCE as the detector. A
492 scanning PSM benefits from its high Λ value while the accuracy is limited by the relatively low size
493 resolution.

494 **Conflicts of interest**

495 There are no conflicts of interest to declare.

496 **Acknowledgement**

497 Financial supports from the National Key R&D Program of China (2017YFC0209503), National
498 Science Foundation of China (21876094 & 41730106), ACTRIS-2 (grant agreement No. 654109),
499 the Academy of Finland (project No. 307331), and Faculty of Science, University of Helsinki, are
500 acknowledged. R. Cai appreciates the support from China Scholarship Council (CSC) for his visit to
501 the University of Helsinki.

502 **References**

- 503 Alanen, J., E. Saukko, K. Lehtoranta, T. Murtonen, H. Timonen, R. Hillamo, P. Karjalainen, H.
504 Kuuluvainen, J. Harra, J. Keskinen, T. Rönkkö. 2015. The formation and physical properties
505 of the particle emissions from a natural gas engine. *Fuel* 162:155-161. doi:
506 10.1016/j.fuel.2015.09.003.
- 507 Almeida, J., S. Schobesberger, A. Kurten, I. K. Ortega, O. Kupiainen-Maatta, A. P. Praplan, A.
508 Adamov, A. Amorim, F. Bianchi, M. Breitenlechner, A. David, J. Dommen, N. M.
509 Donahue, A. Downard, E. Dunne, J. Duplissy, S. Ehrhart, R. C. Flagan, A. Franchin, R.
510 Guida, J. Hakala, A. Hansel, M. Heinritzi, H. Henschel, T. Jokinen, H. Junninen, M. Kajos,
511 J. Kangasluoma, H. Keskinen, A. Kupc, T. Kurten, A. N. Kvashin, A. Laaksonen, K.
512 Lehtipalo, M. Leiminger, J. Leppa, V. Loukonen, V. Makhmutov, S. Mathot, M. J.
513 McGrath, T. Nieminen, T. Olenius, A. Onnela, T. Petaja, F. Riccobono, I. Riipinen, M.
514 Rissanen, L. Rondo, T. Ruuskanen, F. D. Santos, N. Sarnela, S. Schallhart, R. Schnitzhofer,
515 J. H. Seinfeld, M. Simon, M. Sipila, Y. Stozhkov, F. Stratmann, A. Tome, J. Trostl, G.
516 Tsagkogeorgas, P. Vaattovaara, Y. Viisanen, A. Virtanen, A. Vrtala, P. E. Wagner, E.
517 Weingartner, H. Wex, C. Williamson, D. Wimmer, P. Ye, T. Yli-Juuti, K. S. Carslaw, M.
518 Kulmala, J. Curtius, U. Baltensperger, D. R. Worsnop, H. Vehkamäki, J. Kirkby. 2013.

519 Molecular understanding of sulphuric acid-amine particle nucleation in the atmosphere.
520 *Nature* 502:359-363. doi: 10.1038/nature12663.

521 Cai, R., M. Attoui, J. Jiang, F. Korhonen, J. Hao, T. Petäjä, J. Kangasluoma. 2018a.
522 Characterization of a high-resolution supercritical differential mobility analyzer at reduced
523 flow rates. *Aerosol Science and Technology*:online available. doi:
524 10.1080/02786826.2018.1520964.

525 Cai, R., D.-R. Chen, J. Hao, J. Jiang. 2017a. A miniature cylindrical differential mobility analyzer
526 for sub-3 nm particle sizing. *Journal of Aerosol Science* 106:111-119. doi:
527 10.1016/j.jaerosci.2017.01.004.

528 Cai, R., D. Yang, L. R. Ahonen, L. Shi, F. Korhonen, Y. Ma, J. Hao, T. Petäjä, J. Zheng, J.
529 Kangasluoma, J. Jiang. 2018b. Data inversion methods to determine sub-3 nm aerosol size
530 distributions using the particle size magnifier. *Atmospheric Measurement Techniques*
531 11:4477-4491. doi: 10.5194/amt-11-4477-2018.

532 Cai, R., D. Yang, Y. Fu, X. Wang, X. Li, Y. Ma, J. Hao, J. Zheng, J. Jiang. 2017b. Aerosol surface
533 area concentration: A governing factor in new particle formation in beijing. *Atmospheric*
534 *Chemistry and Physics* 17:12327-12340. doi: 10.5194/acp-17-12327-2017.

535 Chen, D.-R., D. Y. H. Pui, D. Hummes, H. Fissan, F. R. Quant, G. J. Sem. 1998. Design and
536 evaluation of a nanometer aerosol differential mobility analyzer (nano-dma). *Journal of*
537 *Aerosol Science* 29:497-509. doi.

538 Chen, M., M. Titcombe, J. Jiang, C. Jen, C. Kuang, M. L. Fischer, F. L. Eisele, J. I. Siepmann, D.
539 R. Hanson, J. Zhao, P. H. McMurry. 2012. Acid-base chemical reaction model for
540 nucleation rates in the polluted atmospheric boundary layer. *Proceedings of the National*
541 *Academy of Sciences of the United States of America* 109:18713-18718. doi:
542 10.1073/pnas.1210285109.

543 Chen, X. and J. Jiang. 2018. Retrieving the ion mobility ratio and aerosol charge fractions for a
544 neutralizer in real-world applications. *Aerosol Science and Technology*:online available. doi:
545 10.1080/02786826.2018.1498587.

546 Enroth, J., J. Kangasluoma, F. Korhonen, S. Hering, D. Picard, G. Lewis, M. Attoui, T. Petäjä.
547 2018. On the time response determination of condensation particle counters. *Aerosol*
548 *Science and Technology* 52:778-787. doi: 10.1080/02786826.2018.1460458.

549 Enting, I. G. and G. N. Newsam. 1990. Atmospheric constituent inversion problems: Implications
550 for baseline monitoring. *Journal of Atmospheric Chemistry* 11:69-87. doi.

551 Fernández de la Mora, J. 2017. Expanded flow rate range of high-resolution nanodmas via
552 improved sample flow injection at the aerosol inlet slit. *Journal of Aerosol Science* 113:265-
553 275. doi: 10.1016/j.jaerosci.2017.07.020.

554 Fernández de la Mora, J. and J. Kozlowski. 2013. Hand-held differential mobility analyzers of high
555 resolution for 1–30nm particles: Design and fabrication considerations. *Journal of Aerosol*
556 *Science* 57:45-53. doi: 10.1016/j.jaerosci.2012.10.009.

557 Fernández de la Mora, J., L. J. Perez-Lorenzo, G. Arranz, M. Amo-Gonzalez, H. Bertscher. 2017.
558 Fast high-resolution nanodma measurements with a 25 ms response time electrometer.
559 *Aerosol Science and Technology* 51:724-734. doi: 10.1080/02786826.2017.1296928.

560 Flagan, R. C. 1999. On differential mobility analyzer resolution. *Aerosol Science and Technology*
561 30:556-570. doi: 10.1080/027868299304417.

562 Hagen, D. E. and D. J. Alofs. 1983. Linear inversion method to obtain aerosol size distributions
563 from measurements with a differential mobility analyzer. *Aerosol Science and Technology*
564 2:465-475. doi.

565 Hoppel, W. A. and G. M. Frick. 1986. Ion—aerosol attachment coefficients and the steady-state
566 charge distribution on aerosols in a bipolar ion environment. *Aerosol Science and*
567 *Technology* 5:1-21. doi: 10.1080/02786828608959073.

568 Iida, K. 2008. Atmospheric nucleation: Development and application of nanoparticle measurements
569 to assess the roles of ion-induced and neutral processes, University of Minnesota,
570 Minneapolis.

571 Iida, K., M. R. Stolzenburg, P. H. McMurry. 2009. Effect of working fluid on sub-2 nm particle
572 detection with a laminar flow ultrafine condensation particle counter. *Aerosol Science and*
573 *Technology* 43:81-96. doi: 10.1080/02786820802488194.

574 Jiang, J., M. Attoui, M. Heim, N. A. Brunelli, P. H. McMurry, G. Kasper, R. C. Flagan, K. Giapis,
575 G. Mouret. 2011a. Transfer functions and penetrations of five differential mobility analyzers
576 for sub-2 nm particle classification. *Aerosol Science and Technology* 45:480-492. doi:
577 10.1080/02786826.2010.546819.

578 Jiang, J., M. Chen, C. Kuang, M. Attoui, P. H. McMurry. 2011b. Electrical mobility spectrometer
579 using a diethylene glycol condensation particle counter for measurement of aerosol size
580 distributions down to 1 nm. *Aerosol Science and Technology* 45:510-521. doi:
581 10.1080/02786826.2010.547538.

582 Jiang, J., C. Kim, X. Wang, M. R. Stolzenburg, S. L. Kaufman, C. Qi, G. J. Sem, H. Sakurai, N.
583 Hama, P. H. McMurry. 2014. Aerosol charge fractions downstream of six bipolar chargers:
584 Effects of ion source, source activity, and flowrate. *Aerosol Science and Technology*
585 48:1207-1216. doi: 10.1080/02786826.2014.976333.

586 Jiang, J., J. Zhao, M. Chen, L. F. Eisele, J. Scheckman, J. B. Williams, C. Kuang, H. P. McMurry.
587 2011c. First measurements of neutral atmospheric cluster and 1-2 nm particle number size
588 distributions during nucleation events. *Aerosol Science and Technology (ARL)* 45:ii-v. doi:
589 10.1080/02786826.2010.546817.

590 Kallinger, P., G. Steiner, W. W. Szymanski. 2012. Characterization of four different bipolar
591 charging devices for nanoparticle charge conditioning. *Journal of Nanoparticle Research*
592 14. doi: 10.1007/s11051-012-0944-z.

593 Kandlikar, M. and G. Ramachandran. 1999. Inversion methods for analyzing aerosol spectrometer
594 measurements: A critical review. *Journal of Aerosol Science* 30:413-437. doi.

595 Kangasluoma, J., L. Ahonen, M. Attoui, H. Vuollekoski, M. Kulmala, T. Petäjä. 2015. Sub-3 nm
596 particle detection with commercial tsi 3772 and airmodus a20 fine condensation particle
597 counters. *Aerosol Science and Technology* 49:674-681. doi:
598 10.1080/02786826.2015.1058481.

599 Kangasluoma, J., L. R. Ahonen, T. M. Laurila, R. Cai, J. Enroth, S. B. Mazon, F. Korhonen, P. P.
600 Aalto, M. Kulmala, M. Attoui, T. Petäjä. 2018. Laboratory verification of a new high flow
601 differential mobility particle sizer, and field measurements in hyytiälä. *Journal of Aerosol*
602 *Science* 124:1-9. doi.

603 Kangasluoma, J., A. Franchin, J. Duplissy, L. Ahonen, F. Korhonen, M. Attoui, J. Mikkilä, K.
604 Lehtipalo, J. Vanhanen, M. Kulmala, T. Petäjä. 2016a. Operation of the airmodus a11 nano
605 condensation nucleus counter at various inlet pressures and various operation temperatures,
606 and design of a new inlet system. *Atmospheric Measurement Techniques* 9:2977-2988. doi:
607 10.5194/amt-9-2977-2016.

608 Kangasluoma, J. and J. Kontkanen. 2017. On the sources of uncertainty in the sub-3 nm particle
609 concentration measurement. *Journal of Aerosol Science* 112:34-51. doi:
610 10.1016/j.jaerosci.2017.07.002.

611 Kangasluoma, J., A. Samodurov, M. Attoui, A. Franchin, H. Junninen, F. Korhonen, T. Kurtén, H.
612 Vehkamäki, M. Sipilä, K. Lehtipalo, D. R. Worsnop, T. Petäjä, M. Kulmala. 2016b.
613 Heterogeneous nucleation onto ions and neutralized ions: Insights into sign-preference. *The*
614 *Journal of Physical Chemistry C* 120:7444-7450. doi: 10.1021/acs.jpcc.6b01779.

615 Kerminen, V.-M., L. Pirjola, M. Kulmala. 2001. How significantly does coagulational scavenging
616 limit atmospheric particle production? *Journal of Geophysical Research: Atmospheres*
617 106:24119-24125. doi: 10.1029/2001jd000322.

Kim, C. S., K. Okuyama, J. Fernández de la Mora. 2003. Performance evaluation of an improved particle size magnifier (psm) for single nanoparticle detection. *Aerosol Science and Technology* 37:791-803. doi: 10.1080/027868203000913.

Kirkby, J., J. Duplissy, K. Sengupta, C. Frege, H. Gordon, C. Williamson, M. Heinritzi, M. Simon, C. Yan, J. Almeida, J. Trostl, T. Nieminen, I. K. Ortega, R. Wagner, A. Adamov, A. Amorim, A. K. Bernhammer, F. Bianchi, M. Breitenlechner, S. Brilke, X. Chen, J. Craven, A. Dias, S. Ehrhart, R. C. Flagan, A. Franchin, C. Fuchs, R. Guida, J. Hakala, C. R. Hoyle, T. Jokinen, H. Junninen, J. Kangasluoma, J. Kim, M. Krapf, A. Kurten, A. Laaksonen, K. Lehtipalo, V. Makhmutov, S. Mathot, U. Molteni, A. Onnela, O. Perakyla, F. Piel, T. Petaja, A. P. Praplan, K. Pringle, A. Rap, N. A. Richards, I. Riipinen, M. P. Rissanen, L. Rondo, N. Sarnela, S. Schobesberger, C. E. Scott, J. H. Seinfeld, M. Sipila, G. Steiner, Y. Stozhkov, F. Stratmann, A. Tome, A. Virtanen, A. L. Vogel, A. C. Wagner, P. E. Wagner, E. Weingartner, D. Wimmer, P. M. Winkler, P. Ye, X. Zhang, A. Hansel, J. Dommen, N. M. Donahue, D. R. Worsnop, U. Baltensperger, M. Kulmala, K. S. Carslaw, J. Curtius. 2016. Ion-induced nucleation of pure biogenic particles. *Nature* 533:521-526. doi: 10.1038/nature17953.

Knutson, E. O. 1976. Extended electric mobility method for measuring aerosol particle size and concentration, in *Fine particles: Aerosol generation, measurement, sampling, and analysis*, Liu, B. Y. H., ed., Academic Press, New York, 739-762.

Kogan, Y. I. and Z. A. Burnasheva. 1960. Growth and measurement of condensation nuclei in a continuous stream. *Zhurnal Fizicheskoi Khimii* 34:2630-2639. doi.

Kuang, C. 2018. A diethylene glycol condensation particle counter for rapid sizing of sub-3 nm atmospheric clusters (online available). *Aerosol Science and Technology*. doi: 10.1080/02786826.2018.1481279.

Kuang, C., M. Chen, J. Zhao, J. Smith, P. H. McMurry, J. Wang. 2012. Size and time-resolved growth rate measurements of 1 to 5 nm freshly formed atmospheric nuclei. *Atmospheric Chemistry and Physics* 12:3573-3589. doi: 10.5194/acp-12-3573-2012.

Kulmala, M., T. Petäjä, T. Nieminen, M. Sipilä, H. E. Manninen, K. Lehtipalo, M. Dal Maso, P. P. Aalto, H. Junninen, P. Paasonen, I. Riipinen, K. E. Lehtinen, A. Laaksonen, V. M. Kerminen. 2012. Measurement of the nucleation of atmospheric aerosol particles. *Nature protocols* 7:1651-1667. doi: 10.1038/nprot.2012.091.

Lee, H. M., C. Soo Kim, M. Shimada, K. Okuyama. 2005. Bipolar diffusion charging for aerosol nanoparticle measurement using a soft x-ray charger. *Journal of Aerosol Science* 36:813-829. doi: 10.1016/j.jaerosci.2004.10.011.

Lehtipalo, K., J. Leppä, J. Kontkanen, J. Kangasluoma, A. Franchin, D. Wimmer, S. Schobesberger, H. Junninen, T. Petäjä, M. Sipilä, J. Mikkilä, J. Vanhanen, D. R. Worsnop, M. Kulmala. 2014. Methods for determining particle size distribution and growth rates between 1 and 3 nm using the particle size magnifier. *Boreal Environment Research* 19:215-236. doi.

Maher, E. F. and N. M. Laird. 1985. Em algorithm reconstruction of particle size distributions from diffusion battery data. *Journal of Aerosol Science* 16:557-570. doi.

Mai, H. and R. C. Flagan. 2018. Scanning dma analysis: I. Classification transfer function. *Aerosol Science and Technology*:online available. doi: 10.1080/02786826.2018.1528005.

Manninen, H. E., T. Petäjä, A. Asmi, I. Riipinen, T. Nieminen, J. Mikkilä, U. Hörrak, A. Mirme, S. Mirme, L. Laakso, V.-M. Kerminen, M. Kulmala. 2009. Long-term field measurements of charged and neutral clusters using neutral cluster and air ion spectrometer (nais). *Boreal Environment Research* 14:591-605. doi.

Markowski, G. R. 1987. Improving twomey's algorithm for inversion of aerosol measurement data. *Aerosol Science and Technology* 7:127-141. doi: 10.1080/02786828708959153.

McMurry, P. H. 1983. New particle formation in the presence of an aerosol: Rates, time scales, and sub-0.01 μm size distributions *Journal of Colloid and Interface Science* 95:72-80. doi.

668 Nosko, O., J. Vanhanen, U. Olofsson. 2016. Emission of 1.3–10 nm airborne particles from brake
 669 materials. *Aerosol Science and Technology* 51:91-96. doi:
 670 10.1080/02786826.2016.1255713.

671 Okuyama, K., Y. Kousaka, T. Motouchi. 1984. Condensational growth of ultrafine aerosol particles
 672 in a new particle size magnifier. *Aerosol Science and Technology* 3:353-366. doi:
 673 10.1080/02786828408959024.

674 Ramachandran, G. and D. W. Cooper, eds. 2011. *Size distribution data analysis and presentation*.
 675 John Wiley & Sons, New York.

676 Russell, L. M., R. C. Flagan, J. H. Seinfeld. 1995. Asymmetric instrument response resulting from
 677 mixing effects in accelerated dma-cpc measurements. *Aerosol Science and Technology*
 678 23:491-509. doi: 10.1080/02786829508965332.

679 Sipilä, M., K. E. Lehtinen, M. Kulmala, T. Petäjä, H. Junninen, P. Aalto, H. E. Manninen, E.-M.
 680 Kyrö, A. Asmi, I. Riipinen, J. Curtius, A. Kürten, S. Borrmann, C. D. O'Dowd. 2008.
 681 Applicability of condensation particle counters to measure atmospheric clusters.
 682 *Atmospheric Chemistry and Physics* 8:4049-4060. doi.

683 Sipilä, M., K. Lehtipalo, M. Attoui, K. Neitola, T. Petäjä, P. P. Aalto, C. D. O'Dowd, M. Kulmala.
 684 2009. Laboratory verification of ph-cpc's ability to monitor atmospheric sub-3 nm clusters.
 685 *Aerosol Science and Technology* 43:126-135. doi: 10.1080/02786820802506227.

686 Sipilä, M., N. Sarnela, T. Jokinen, H. Henschel, H. Junninen, J. Kontkanen, S. Richters, J.
 687 Kangasluoma, A. Franchin, O. Peräkylä, M. P. Rissanen, M. Ehn, H. Vehkamäki, T. Kurten,
 688 T. Berndt, T. Petäjä, D. Worsnop, D. Ceburnis, V. M. Kerminen, M. Kulmala, C. O'Dowd.
 689 2016. Molecular-scale evidence of aerosol particle formation via sequential addition of hio3.
 690 *Nature* 537:532-534. doi: 10.1038/nature19314.

691 Stolzenburg, D., G. Steiner, P. M. Winkler. 2017. A dma-train for precision measurement of sub-
 692 10 nm aerosol dynamics. *Atmospheric Measurement Techniques* 10:1639-1651. doi:
 693 10.5194/amt-10-1639-2017.

694 Stolzenburg, M. R. 2018. A review of transfer theory and characterization of measured performance
 695 for differential mobility analyzers. *Aerosol Science and Technology*:online available. doi:
 696 10.1080/02786826.2018.1514101.

697 Stolzenburg, M. R. and H. P. McMurry. 1991. An ultrafine aerosol condensation nucleus counter.
 698 *Aerosol Science and Technology* 14:48-65. doi.

699 Stolzenburg, M. R. and P. H. McMurry. 2008. Equations governing single and tandem dma
 700 configurations and a new lognormal approximation to the transfer function. *Aerosol Science*
 701 *and Technology* 42:421-432. doi: 10.1080/02786820802157823.

702 Stolzenburg, M. R., J. H. T. Scheckman, M. Attoui, H.-S. Han, P. H. McMurry. 2018.
 703 Characterization of the tsi model 3086 differential mobility analyzer for classifying aerosols
 704 down to 1 nm. *Aerosol Science and Technology* 52:748-756. doi:
 705 10.1080/02786826.2018.1456649.

706 Tammet, H. 2011. Symmetric inclined grid mobility analyzer for the measurement of charged
 707 clusters and fine nanoparticles in atmospheric air. *Aerosol Science and Technology* 45:468-
 708 479. doi: 10.1080/02786826.2010.546818.

709 Tang, Q., R. Cai, X. You, J. Jiang. 2017. Nascent soot particle size distributions down to 1 nm from
 710 a laminar premixed burner-stabilized stagnation ethylene flame. *Proceedings of the*
 711 *Combustion Institute* 36:993-1000. doi: 10.1016/j.proci.2016.08.085.

712 Tröstl, J., T. Tritscher, O. F. Bischof, H.-G. Horn, T. Krinke, U. Baltensperger, M. Gysel. 2015.
 713 Fast and precise measurement in the sub-20nm size range using a scanning mobility particle
 714 sizer. *Journal of Aerosol Science* 87:75-87. doi: 10.1016/j.jaerosci.2015.04.001.

715 Vanhanen, J., J. Mikkilä, K. Lehtipalo, M. Sipilä, H. E. Manninen, E. Siivola, T. Petäjä, M.
 716 Kulmala. 2011. Particle size magnifier for nano-cn detection. *Aerosol Science and*
 717 *Technology* 45:533-542. doi: 10.1080/02786826.2010.547889.

718 Wang, J., R. C. Flagan, J. H. Seinfeld. 2002. Diffusional losses in particle sampling systems
 719 containing bends and elbows. *Journal of Aerosol Science* 33:843/857. doi.
 720 Wang, S. C. and R. C. Flagan. 1990. Scanning electrical mobility spectrometer. *Aerosol Science*
 721 *and Technology* 13:230-240. doi: 10.1080/02786829008959441.
 722 Weber, R. J., M. R. Stolzenburg, S. N. Pandis, H. P. McMurry. 1998. Inversion of ultrafine
 723 condensation nucleus counter pulse height distributions to obtain nanoparticle (~3-10 nm)
 724 size distributions. *Journal of Aerosol Science* 29:601-615. doi.
 725 Wiedensohler, A. 1988. An approximation of the bipolar charge distribution for particles in the
 726 submicron size range. *Journal of Aerosol Science* 19:387-389. doi.
 727 Wiedensohler, A., W. Birmili, A. Nowak, A. Sonntag, K. Weinhold, M. Merkel, B. Wehner, T.
 728 Tuch, S. Pfeifer, M. Fiebig, A. M. Fjåraa, E. Asmi, K. Sellegri, R. Depuy, H. Venzac, P.
 729 Villani, P. Laj, P. Aalto, J. A. Ogren, E. Swietlicki, P. Williams, P. Roldin, P. Quincey, C.
 730 Hüglin, R. Fierz-Schmidhauser, M. Gysel, E. Weingartner, F. Riccobono, S. Santos, C.
 731 Grüning, K. Faloon, D. Beddows, R. Harrison, C. Monahan, S. G. Jennings, C. D. O'Dowd,
 732 A. Marinoni, H. G. Horn, L. Keck, J. Jiang, J. Scheckman, P. H. McMurry, Z. Deng, C. S.
 733 Zhao, M. Moerman, B. Henzing, G. de Leeuw, G. Löschau, S. Bastian. 2012. Mobility
 734 particle size spectrometers: Harmonization of technical standards and data structure to
 735 facilitate high quality long-term observations of atmospheric particle number size
 736 distributions. *Atmospheric Measurement Techniques* 5:657-685. doi: 10.5194/amt-5-657-
 737 2012.
 738 Zhang, R., A. Khalizov, L. Wang, M. Hu, W. Xu. 2012. Nucleation and growth of nanoparticles in
 739 the atmosphere. *Chemical Reviews* 112:1957-2011. doi: 10.1021/cr2001756.
 740

741 Table 1. Specifications of the four electrical mobility spectrometers tested in this study.

Instrument	DMA model	aerosol-to-sheath flow ratio (lpm/lpm)	Detector model	t (s)	Q (cm ³ /s)
DEG-SMPS 2011	TSI nanoDMA	2/15	DEG-UCPC + TSI 3772 CPC	20	0.60
DEG-SMPS 2017	Mini-cyDMA	2.5/25	DEG-UCPC + TSI 3772 CPC	8.8	4.17
Half-mini DMPS	Half-mini DMA	5/100	Airmodus A10 PSM + A20 CPC	10	11.4
DMA-train	GRIMM S-DMA	2.5/25	Airmodus A10 PSM + TSI 3776 UCPC	300*	0.83

742 *: The averaging time of the DMA train is assumed to be 300 s in this study for the same temporal resolution to the other three
743 instruments.

744

Figure 1. (a) The overall detection efficiencies, P , (b) the II values, (c) the expected values of particle counts, and (d) the relative standard deviation of particle counts as a function of particle electrical mobility diameters. The aerosol size distribution function, $dN/d\log d_p$, was assumed to be a constant value of $5 \times 10^5 \text{ cm}^{-3}$ when estimating the particle raw count number.

Figure 2. The relationship between the relative statistical uncertainty and Π at different $dN/d\log d_p$ levels. Uncertainties larger than 100 percent are shown in dashed lines because the corresponding expected values of particle raw count numbers are smaller than one.

Figure 3. (a) The relationship between the limit of detection and the II parameter at different false count levels. The limit of detection is usually determined using one false count, whereas larger false count numbers account for the conditions that there are higher background noises due to homogeneous nucleation of the working fluid and/or unstable particle sources. (b) The estimated limits of detection of different instruments. The limits of detection are determined as the corresponding $dN/d\log d_p$ values of one single CPC count or three times the root mean square of the FCE noise during the measurement time of each size bin. The representative diameters for the NAIS are the peak diameters of the kernel functions for each size bin. The NAIS data points with diameters smaller than 2 nm are shown in open circles because that high concentration of ion clusters is often detected in this size range during atmospheric measurements and they are not necessarily related to atmospheric new particle formation.

769

770 Figure 4. The overall detection efficiency, P , and the Π value for 1.2 nm negatively charged tungsten oxide particles
 771 as a function of the CPC inlet flow rate. The CPC is assumed to be used as the particle detector of the half-mini
 772 DMPS for ease of illustration. It is a butanol CPC working at a relatively high background noise level due to the
 773 homogeneous nucleation of butanol. The CPC detection efficiencies are adapted from Kangasluoma et al. (2015).
 774 Other parameters for estimating P and Π are assumed to be independent of the CPC flow rate, e.g., the sampling
 775 flow rate and the aerosol flow rate for the DMA at different CPC flow rates are assumed to be a constant value of 5
 776 lpm.

777

778

779

780

781

782

783

784

Figure 5. The test size distribution and statistical uncertainties at the Π values of the DEG-SMPS 2011 and the DEG-SMPS 2017. A large Π value 10 times that of the DEG-SMPS 2017 is also included. The effect of size resolution on the inverted particle raw counts is neglected. The Poisson distribution is approximated by a normal distribution. A set of simulation results are shown in scatters as examples. The boundaries for 95% confidence intervals of the inverted results are indicated.

Figure 6. The relationship between the relative error of the inverted aerosol size distribution and the size resolution. The size resolutions in the legends indicate the DMA resolutions for 1.5 nm particles. The resolutions for other electrical mobility diameters are estimated using Eqs. 6, 7, and 8. The test size distribution is the same as the distribution in Fig. 5. (a) The inverted aerosol size distributions at different size resolutions. The shaded areas represent the random errors due to statistical uncertainties estimated using a II value 10 times that of DEG-SMPS 2017. The 95% confidence intervals are estimated using the normal distribution. Note the probabilities of the inverted $dN/d\log d_p$ values are not evenly distributed in the shaded areas. (b) The relative systematic errors of the inverted aerosol size distribution at different diameters as a function of the DMA resolution. (c) The relative uncertainties of two electrical mobility spectrometers with different size resolutions and a scanning particle size magnifier. The inversion method for the particle size magnifier is a stepwise method used in commercial instruments (Lehtipalo et al. 2014).

Figure 7. (a) The estimated uncertainties of the measured aerosol size distribution as a function of II and the size resolution. The value of uncertainty is indicated by color. The $dN/d\log d_p$ is equal to that at 1.5 nm of the test aerosol size distribution in Figs. 5 and 6 ($4.9 \times 10^5 \text{ cm}^{-3}$). The estimated uncertainties at a lower $dN/d\log d_p$ value ($1 \times 10^5 \text{ cm}^{-3}$) and a higher $dN/d\log d_p$ value ($3 \times 10^6 \text{ cm}^{-3}$) are shown in Fig S4. The size resolutions of the four instruments shown in open circles are the size resolutions for 1.5 nm particles. The variations of II , DMA resolution, and the corresponding uncertainty as a function of the DMA sheath flow rate while maintaining other configurations are indicated using dashed lines. (b) The uncertainty for the DEG-SMPS 2017 as a function of DMA resolution when varying the sheath flow rate, corresponding the dashed line in Fig. 7a. The total uncertainty is characterized by the sum of the systematic error and the random error.